

Top gasket 1/2" x 1/2"
 adhesive backed sponge
 rubber seal
 (McMaster-Carr
 #93085K13)

3/4" plastic laminated fiberboard

2.25"

PVC wrapped 1" fiberglass
 acoustical panel from
 Acoustical Solutions, held
 in place by a 27" tall
 piece of limp mass
 material (AB10NR), also
 from Acoustical Solutions

31.75"

Side gasket 3/4" x 3/8"
 adhesive backed sponge
 rubber seal
 (McMaster-Carr
 #93085K15)


33.43"

3/4" plastic laminated fiberboard


79°

8"


1/2" vinyl coated sponge gasket,
 (McMaster-Carr #1141A4), fastened
 with #8x5/8" oval phillips screws
 (McMaster-Carr #90398A196)




QTY REQD	FSCM NO.	PART OR IDENTIFYING NO.	NOMENCLATURE OR DESCRIPTION	MATERIAL SPECIFICATION
PARTS LIST				
UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES TOLERANCES ARE: FRACTIONS DECIMALS ANGLES ± .XX±.01 ± .XXX±.005 .XXXX±.0005		CONTRACT NO.		<i>LIGO Project</i> <i>California Institute of Technology</i> <i>Massachusetts Institute of Technology</i>
MATERIAL Per Plan		APPROVALS DRAWN T. Mahood	DATE 02/08/01	<i>PSL Pre-stabilized Laser</i> Acoustic Skirt Cross section and layout
FINISH		CHECKED	ISSUED	SIZE B
				FSCM NO.
				DWG. NO. D010051-00-W
				REV.
				SCALE
				SHEET 1 of 2



Std. Side Panels - 14 Required





Notched Side Panel - 1 Required



Notched Side Panel - 1 Required

- Notes:
1. All material to be 3/4" thick plastic laminate fiberboard or plywood.
 2. Laminate color to be white.
 3. All edges to be laminated. No exposed wood.
 4. Bottom joint to be doweled.
 5. All surfaces are square to each other unless otherwise noted.



Corner Panels - 4 Required

QTY REQD	FSCM NO.	PART OR IDENTIFYING NO.	NOMENCLATURE OR DESCRIPTION	MATERIAL SPECIFICATION
PARTS LIST				
UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES TOLERANCES ARE: FRACTIONS DECIMALS ANGLES ± .XX±.01 ± .XXX±.005 .XXXX±.0005		CONTRACT NO.		<i>LIGO Project</i> <i>California Institute of Technology</i> <i>Massachusetts Institute of Technology</i>
MATERIAL Per Plan		APPROVALS T. Mahood	DATE 02/08/01	PSL Pre-stabilized Laser
NEXT ASSY	USED ON	FINISH	ISSUED	SIZE B FSCM NO. DWG. NO. D010051-00-W REV.
APPLICATION		DO NOT SCALE DRAWING		SCALE SHEET 2 of 2