




NOTES: (UNLESS OTHERWISE SPECIFIED)
PART MUST BE FREE OF PAINT AND MILL MARKS

REV.	DATE	DCN #	DRAWING TREE #
A	3.31.03	E030164-00-D	-

DRILL 0.177 DIA THRU, C'TR
BORE 0.313 DIA TO DEPTH
SHOWN

DRILL 0.089 THRU
TAP 4-40 UNC-2B X 0.75 DP,
2 PLCS



ITEM NO.	REQ.	SPARE	TOT. QUANTITY	PART NUMBER	DESCRIPTION	SUPPLIER	
PARTS LIST							
UNLESS OTHERWISE SPECIFIED: DIMENSIONS ARE IN INCHES				DATE	NAME	 CALIFORNIA INSTITUTE OF TECHNOLOGY MASSACHUSETTS INSTITUTE OF TECHNOLOGY	
TOLERANCES: .XX ± 0.01 .XXX ± 0.005				DRAWN	3.31.03		M.R. SMITH
ANGULAR: X ± 1 DEG XX ± 0.2 DEG				CHECKED			
MATERIAL: ALUM 6061-T6				COMMENTS: BREAK CORNERS AND EDGES REMOVE ALL BURRS		40 METER INTERFEROMETER	
FINISH: NATURAL						FARADAY ISOLATOR	
						PLATFORM-BREWSTER MIRROR	
SIZE B		DWG. NO. D030175-A-D			REV A		
SCALE:					SHEET 1 OF 1		