


Fine Adjust = +/- 6 mV
Coarse Adjust = +/- 60 mV

G=400

G = 4 w/ pole @ 800Hz or
G = 20 w/ pole @ 160Hz



Title			LIGO Laboratory California Institute of Technology Massachusetts Institute of Technology			LIGO 		
PZT Strain Gauge Preamplifier								
Size: B	DCC Number: D060285	SCH / PCB Revision: A	Engineer: Joshua R. Myers	Date: 12/14/2006				
				Time: 8:01:01 AM				
File: C:\Documents and Settings\jmyers\My Documents\LHOBoards\Joshua\PZT Strain Gauge Preamp\PZT Strain Gauge1.SchDoc								Sheet 2 of 2