

Discrepancy Between Photon Calibration and Official Calibration

Peter Kalmus
Columbia University


on behalf of the
Calibration Committee

Commissioning Meeting
May 15, 2006



Mike Landry and Rick Savage's First Look


Biggest discrepancy is 7%.
We haven't done so well since.





2k shows discrepancy may be increasing with frequency

4k propagation might not be valid

alpha and beta were not available at the time this was done – they were extrapolated



LIGO Recent Measurements with DMT



1000 second integrations, measurements of peak only

H1X mean is 1.4x reference; H1Y is 1.6x reference

H2 means are both 1.6x reference

it's really ~15% worse than this: noise power has not been subtracted from DARM_ERR peak measurement; this will shift PCal higher

standard deviations between 7% and 10%

Get a DMT started at Livingston

Subtract power from noise in DMT

use nearest neighbor bins in DMT (tests in Matlab on sandbox data imply that this probably won't change things significantly)

Propagate DARM response function with FDCalibrate in DMT

Test hypothesis that discrepancy increases with frequency

try adding a pcal line with H1X for one week at a lower frequency.

719.1 Hz should be clear of pulsars, violin modes, and pulsar injections

804.85 Hz ($1609.7 / 2$) is clear as well

probably OK to keep drive level the same....

Inject a line (on an ETM) with the photon cal laser / AOM

Measure peak in DARM_ERR PSD

- use 3 bins (peak + nearest neighbors)

- subtract power from noise, estimated from nearby bins (or use fit)

- take sqrt to get peak amplitude

Measure peak in ETMX_CAL PSD

- bin width and windowing function factors cancel

Convert ETMX_CAL measurement to strain

- readout cts \rightarrow Watts out of enclosure (at DC)

- account for power lost due to reflection off of viewport

- account for momentum lost due to transmission through optic

- measure angle of incidence, mass of optic

- systematic errors: off-centered beam in H1 (~3%); AOM to photodetector transfer function (~2%); power meter (up to 10%)

Power meter systematic – readings possibly ~10% high
would cause cal factor to be ~10% too high
currently under investigation – meters have been sent to Scientech for comparison

We assume test mass displacement is given by:

$$x_{exc}(\omega) = \frac{2P \cos \theta}{Mc \omega^2}$$

Heating effect would change f^2 dependency, but this was checked

Would expect this to be worse at lower frequencies

Transfer function ETMY_CAL/ ETMY_CAL_EXC attenuates by 2% from DC at 1600 Hz
still a mystery

would cause the cal factor to be lower, not higher

Blunder

sqrt(2) somewhere? Not correcting for windowing function?

hmm.

Not using the correct reference response function

Photodetector calibrations may have drifted

Harmonics

They account for less than 1% of power; also, they will “factor out” of the result

Absorption in viewport

Other means by which power is not getting to TM, or TM is not moving as much as we think.